

CHLAMYDIA
TRACHOMATIS
NUCLEIC ACID
AMPLIFICATION
TEST – Urine

Orderable - NGCAD

Turnaround Time: Up to 3 days

Specimen Type and Collection Information:

Please review the Public Health Ontario Laboratory Test Information Index:

[Chlamydia trachomatis/Neisseria gonorrhoeae \(CT/GC\) – Nucleic Acid Amplification Testing \(NAAT\) - Urine](#)



Laboratory:
Microbiology



Requisition:
[PUBLIC HEALTH
LABORATORY TEST
REQUISITION](#)

This requisition **MUST**
accompany the
specimen



Method of Analysis:
Nucleic Acid Amplification
(NAAT)



Test Schedule:
Referred out weekdays



Referred to:
Public Health Laboratory
1200 Commissioners Rd E
London, Ontario
519-455-9310

Specimen Type	Minimum Volume	Collection Kit
Urine (first catch)	2.0 mL	Transfer 2 ml of first catch urine from a sterile container to the Hologic Aptima® Urine Transport tube between the black fill lines using the disposable pipette within 24 hours of collection.

Note: the same specimen can be tested for Neisseria gonorrhoeae NAAT.



Storage and Shipment:

Samples can be stored at room temperature or refrigerated. **DO NOT FREEZE.**
The specimen must be received within 24 hours of collection.
Do **NOT** break foil on top of cap.