

EVAC TUBES

RRTs will confirm patency of lumen and document Q2-Q3H.

Must use EVAC specific suction regulator (-20 mmHg).

Occasionally, a properly functioning EVAC tube will imitate the sound of a cuff leak.

Rule out this suction sound prior to adding additional air to the cuff.

Disconnect and cap prior to patient turns or transports.

